

CHANGES TO AATB STANDARDS FOR TISSUE BANKING

SECTION E - PROCESSING, PRESERVATION, QUARANTINE, AND STORAGE

E1.000 PROCESSING, PRESERVATION, QUARANTINE, AND STORAGE— GENERAL

Current (13th edition; refer to update posted July 11, 2013 (AATB Bulletin No. 13-13))

E1.300 Reagents, Supplies, Materials, and Equipment

(C, V, MS, OA, S)

All non-disposable surgical instruments and mechanical/electrical equipment used in tissue *Processing* shall be cleaned, disinfected, and *Sterilized* between use for tissue from different donors according to written procedures. For non-disposable surgical instruments and mechanical/electrical equipment deemed *Critical*, written procedures must be prepared and should be *Validated* to prevent contamination or *Cross-Contamination* during *Processing*.

(with amendments)

E1.300 Reagents, Supplies, Materials, and Equipment

(C, V, MS, OA, S, CT)

All non-disposable surgical instruments and mechanical/electrical equipment used in tissue *Processing* shall be cleaned, disinfected, and *Sterilized* between use for tissue from different donors according to written procedures. For non-disposable surgical instruments and mechanical/electrical equipment deemed *Critical*, written procedures must be prepared and should be *Validated* to prevent contamination or *Cross-Contamination* during *Processing*.

(as amended)

E1.300 Reagents, Supplies, Materials, and Equipment

(C, V, MS, OA, S, CT)

All non-disposable surgical instruments and mechanical/electrical equipment used in tissue *Processing* shall be cleaned, disinfected, and *Sterilized* between use for tissue from different donors according to written procedures. For non-disposable surgical instruments and mechanical/electrical equipment deemed *Critical*, written procedures must be prepared and should be *Validated* to prevent contamination or *Cross-Contamination* during *Processing*.

Publication date: November 12, 2013 (AATB Bulletin No. 13-18)

Effective date: February 10, 2014 (in 90 days)